

Biological evaluation of medical devices —

Part 1: Evaluation and testing

The European Standard EN ISO 10993-1:2003 has the status of a
British Standard

ICS 11.100

National foreword

This British Standard is the official English language version of EN ISO 10993-1:2003. It supersedes BS EN ISO 10993-1:1998 which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee CH/194, Biological evaluation of medical devices, which has the responsibility to:

- aid enquirers to understand the text;
- present to the responsible international/European committee any enquiries on the interpretation, or proposals for change, and keep the UK interests informed;
- monitor related international and European developments and promulgate them in the UK.

A list of organizations represented on this committee can be obtained on request to its secretary.

Additional information

Users of this standard should note that ISO/TC 194, Biological evaluation of medical devices, originally initiated this revision of EN ISO 10993-1 only for the purpose of user clarification and not to alter its technical provisions in any way. However, additional crosses were introduced into Table 1 and Table 2, which represented a significant technical change to the standard. The UK was against this change and accordingly voted against this standard at the Formal Vote stage.

Cross-references

The British Standards which implement international or European publications referred to in this document may be found in the *BSI Catalogue* under the section entitled "International Standards Correspondence Index", or by using the "Search" facility of the *BSI Electronic Catalogue* or of British Standards Online.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

Compliance with a British Standard does not of itself confer immunity from legal obligations.

Summary of pages

This document comprises a front cover, an inside front cover, the EN ISO title page, the EN ISO foreword page, the ISO title page, pages ii to vi, pages 1 to 14, the Annex ZA page and a back cover.

The BSI copyright notice displayed in this document indicates when the document was last issued.

Amendments issued since publication

Amd. No.	Date	Comments

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 9 October 2003

© BSI 9 October 2003

ISBN 0 580 42731 5

English version

Biological evaluation of medical devices - Part 1: Evaluation and testing (ISO 10993-1:2003)

Evaluation biologique des dispositifs médicaux - Partie 1:
Evaluation et essais (ISO 10993-1:2003)

Biologische Beurteilung von Medizinprodukten Teil 1:
Beurteilung und Prüfungen (ISO 10993-1:2003)

This European Standard was approved by CEN on 28 July 2003.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Luxembourg, Malta, Netherlands, Norway, Portugal, Slovakia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

CORRECTED 2003-09-24

Foreword

This document (EN ISO 10993-1:2003) has been prepared by Technical Committee ISO/TC 194 "Biological evaluation of medical devices" in collaboration with Technical Committee CEN/TC 206 "Biocompatibility of medical and dental materials and devices", the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by February 2004, and conflicting national standards shall be withdrawn at the latest by February 2004.

This document supersedes EN ISO 10993-1:1997.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA, which is an integral part of this document.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Luxembourg, Malta, Netherlands, Norway, Portugal, Slovakia, Spain, Sweden, Switzerland and the United Kingdom.

Endorsement notice

The text of ISO 10993-1:2003 has been approved by CEN as EN ISO 10993-1:2003 without any modifications.

INTERNATIONAL
STANDARD

ISO
10993-1

Third edition
2003-08-01

**Biological evaluation of medical
devices —**

**Part 1:
Evaluation and testing**

Évaluation biologique des dispositifs médicaux —

Partie 1: Évaluation et essais



Reference number
ISO 10993-1:2003(E)

EN ISO 10993-1:2003

Contents

Page

Foreword	iv
Introduction	vi
1 Scope	1
2 Terms and definitions	1
3 General principles applying to biological evaluation of medical devices	2
4 Categorization of medical devices	3
5 Testing	4
6 Selection of biological evaluation tests	7
7 Assurance of test methods	8
Annex A (informative) Rationale	11
Annex B (informative) Flow chart to aid in ensuring a systematic approach to biological evaluation of medical devices	13
Bibliography	14

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 10993-1 was prepared by Technical Committee ISO/TC 194, *Biological evaluation of medical devices*.

This third edition cancels and replaces the second edition (ISO 10993-1:1997), of which it constitutes a minor revision.

ISO 10993 consists of the following parts, under the general title *Biological evaluation of medical devices*:

- *Part 1: Evaluation and testing*
- *Part 2: Animal welfare requirements*
- *Part 3: Tests for genotoxicity, carcinogenicity and reproductive toxicity*
- *Part 4: Selection of tests for interactions with blood*
- *Part 5: Tests for in vitro cytotoxicity*
- *Part 6: Tests for local effects after implantation*
- *Part 7: Ethylene oxide sterilization residuals*
- *Part 8: Selection and qualification of reference materials for biological tests*
- *Part 9: Framework for identification and quantification of potential degradation products*
- *Part 10: Tests for irritation and delayed-type hypersensitivity*
- *Part 11: Tests for systemic toxicity*
- *Part 12: Sample preparation and reference materials*
- *Part 13: Identification and quantification of degradation products from polymeric medical devices*
- *Part 14: Identification and quantification of degradation products from ceramics*
- *Part 15: Identification and quantification of degradation products from metals and alloys*
- *Part 16: Toxicokinetic study design for degradation products and leachables*

- *Part 17: Establishment of allowable limits for leachable substances*
- *Part 18: Chemical characterization of materials*

Future parts will deal with other relevant aspects of biological testing.

Introduction

This part of ISO 10993 is a combination/harmonization of numerous international and national standards and guidelines concerning the biological evaluation of medical devices. It is intended to be the overall guidance document for the selection of tests enabling evaluation of biological responses relevant to the safety of medical devices and materials.

The role of this part of ISO 10993 is to serve as a framework in which to plan such a biological evaluation which minimizes the number and exposure of test animals.

The protection of humans is the primary goal of ISO 10993.

The appropriate selection and interpretation of biological evaluation tests requires an understanding of the rationale behind such testing. An informative rationale for the use of this part of ISO 10993 is provided in Annex A. Annex B contains a flow chart to aid in the systematic approach to the biological evaluation of medical devices. A bibliography is given at the end of the text.

Biological evaluation of medical devices —

Part 1: Evaluation and testing

1 Scope

This part of ISO 10993 describes

- a) the general principles governing the biological evaluation of medical devices;
- b) the categorization of devices based on the nature and duration of their contact with the body;
- c) the selection of appropriate tests.

This part of ISO 10993 does not cover testing of materials and devices that do not come into direct or indirect contact with the patient's body, nor does it cover biological hazards arising from any mechanical failure.

NOTE Other parts of ISO 10993 cover specific tests (see also the rationale in A.2).

2 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

2.1

medical device

any instrument, apparatus, appliance, material or other article, including software, whether used alone or in combination, intended by the manufacturer to be used for human beings solely or principally for the purpose of:

- diagnosis, prevention, monitoring, treatment or alleviation of disease;
- diagnosis, monitoring, treatment, alleviation of or compensation for an injury or handicap;
- investigation, replacement or modification of the anatomy or of a physiological process;
- control of conception;

and which does not achieve its principal intended action in or on the human body by pharmacological, immunological or metabolic means, but which may be assisted in its function by such means

NOTE 1 Devices are different from drugs, and their biological evaluation requires a different approach.

NOTE 2 Use of the term "medical device" includes dental devices.

2.2

material

any synthetic or natural polymer, metal, alloy, ceramic or other nonviable substance, including tissue rendered nonviable, used as a medical device or any part thereof

2.3

final product

medical device in its "as-used" state

3 General principles applying to biological evaluation of medical devices

3.1 The selection and evaluation of any material or device intended for use in humans requires a structured programme of assessment.

In the design process, an informed decision shall be made and documented that weighs the advantages/disadvantages of the various choices of material and test procedure. To give assurance that the final product will perform as intended and be safe for human use, the programme shall include a biological evaluation.

The biological evaluation shall be planned, carried out and documented by knowledgeable and experienced individuals capable of making informed decisions based on the advantages and disadvantages of the various materials and test procedures available.

3.2 In the selection of materials to be used in device manufacture, the first consideration should be fitness for purpose with regard to characteristics and properties of the material, which include chemical, toxicological, physical, electrical, morphological and mechanical properties.

3.3 The following should be considered for their relevance to the overall biological evaluation of the device:

- a) the material(s) of manufacture;
- b) intended additives, process contaminants and residues;
- c) leachable substances;
- d) degradation products;
- e) other components and their interactions in the final product;
- f) the properties and characteristics of the final product.

NOTE If appropriate, identification and quantification of extractable chemical entities of the final product should precede biological evaluation (see ISO 10993-9).

3.4 Tests to be used in biological evaluation, and the interpretation of the results of such tests, should take into account the chemical composition of the materials, including the conditions of exposure and the nature, degree, frequency and duration of exposure of the device or its constituents to the body. By following these principles, devices can be categorized to facilitate the selection of appropriate tests (see Clause 4). This part of ISO 10993 is concerned with the tests to be carried out on materials and/or the final product.

The range of potential biological hazards is wide and may include:

- a) short-term effects (e.g. acute toxicity, irritation to the skin, eye and mucosal surfaces, sensitization, haemolysis and thrombogenicity);
- b) long-term or specific toxic effects [e.g. subchronic and chronic toxic effects, sensitization, genotoxicity, carcinogenicity (tumorigenicity) and effects on reproduction including teratogenicity].

3.5 All potential biological hazards should be considered for every material and final product, but this does not imply that testing for all potential hazards will be necessary or practical (see Clause 6).

3.6 Any *in vitro* or *in vivo* tests shall be based on end-use applications and appropriate good laboratory practice followed by evaluation by competent informed persons. Whenever possible, *in vitro* screening should

be carried out before *in vivo* tests are commenced. Test data, complete to the extent that an independent analysis could be made, shall be retained (see A.2, "Subclause 3.6").

3.7 The materials or final product shall be considered for biological re-evaluation if any of the following occurs:

- a) any change in the source or in the specification of the materials used in the manufacture of the product;
- b) any change in the formulation, processing, primary packaging or sterilization of the product;
- c) any change in the final product during storage;
- d) any change in the intended use of the product;
- e) any evidence that the product may produce adverse effects when used in humans.

3.8 The biological evaluation performed in accordance with this part of ISO 10993 should be considered in conjunction with the nature and mobility of the ingredients in the materials used to manufacture the device and other information, other non-clinical tests, clinical studies and post-market experience for an overall assessment (see A.2, "Subclause 3.8").

4 Categorization of medical devices

4.1 General

Following the general principles laid down in Clause 3, medical devices can be categorized to facilitate the selection of appropriate tests.

The testing of any device that does not fall into one of the categories described should follow the general principles contained in this part of ISO 10993. Certain devices may fall into more than one category, in which case testing appropriate to each category should be considered.

Medical devices shall be categorized according to the nature and duration of body contact as described in 4.2 and 4.3.

4.2 Categorization by nature of body contact

4.2.1 Non-contact devices

Medical devices that do not contact the patient's body directly or indirectly are not included in the scope of ISO 10993.

4.2.2 Surface-contacting devices

These include medical devices in contact with the following surfaces:

- a) **skin:** devices that contact intact skin surfaces only; examples include electrodes, external prostheses, fixation tapes, compression bandages and monitors of various types;
- b) **mucosal membranes:** devices that contact intact mucosal membranes; examples include contact lenses, urinary catheters, intravaginal and intrainestinal devices (stomach tubes, sigmoidoscopes, colonoscopes, gastroscopes), endotracheal tubes, bronchoscopes, dental prostheses, orthodontic devices and intrauterine devices;
- c) **breached or compromised surfaces:** devices that contact breached or otherwise compromised body surfaces; examples include dressings, healing devices and occlusive patches for ulcers, burns and granulation tissue.

4.2.3 External communicating devices

These include medical devices in contact with the following application sites:

- a) **blood path, indirect:** devices that contact the blood path at one point and serve as a conduit for entry into the vascular system; examples include solution administration sets, extension sets, transfer sets and blood administration sets;
- b) **tissue/bone/dentin:** devices that contact tissue, bone or pulp/dentin systems; examples include laparoscopes, arthroscopes, draining systems, dental cements, dental filling materials and skin staples;
- c) **circulating blood:** devices that contact circulating blood; examples include intravascular catheters, temporary pacemaker electrodes, oxygenators, extracorporeal oxygenator tubing and accessories, dialysers, dialysis tubing and accessories, haemoadsorbents and immunoadsorbents.

4.2.4 Implant devices

These include medical devices in contact with the following application sites:

- a) **tissue/bone:**
 - 1) devices principally contacting bone; examples include orthopaedic pins, plates, replacement joints, bone prostheses, bone cements and intraosseous devices;
 - 2) devices principally contacting tissue and tissue fluid; examples include pacemakers, drug supply devices, neuromuscular sensors and stimulators, replacement tendons, breast implants, artificial larynxes, subperiosteal implants and ligation clips;
- b) **blood:** devices principally contacting blood; examples include pacemaker electrodes, artificial arteriovenous fistulae, heart valves, vascular grafts, internal drug-delivery catheters and ventricular assist devices.

4.3 Categorization by duration of contact

Medical devices shall be categorized according to the duration of contact as follows:

- a) **Limited exposure (A):** devices whose single or multiple use or contact is likely to be up to 24 h;
- b) **Prolonged exposure (B):** devices whose single, multiple or long-term use or contact is likely to exceed 24 h but not 30 days;
- c) **Permanent contact (C):** devices whose single, multiple or long-term use or contact exceeds 30 days.

If a material or device may be placed in more than one duration category, the more rigorous testing requirements shall apply. With multiple exposures to the device, the decision into which category a device is placed should take into account the potential cumulative effect, bearing in mind the period of time over which these exposures occur.

5 Testing

5.1 General

In addition to the general principles laid down in Clause 3, the following shall apply to biological testing of medical devices.

- a) Testing shall be performed on the final product, or on representative samples taken from the final product or from materials processed in the same manner as the final product.

- b) The choice of test procedures shall take into account:
- 1) the nature, degree, duration, frequency and conditions of exposure to or contact of humans with the device in the normal intended use;
 - 2) the chemical and physical nature of the final product;
 - 3) the toxicological activity of the chemical elements or compounds in the formulation of the final product;
 - 4) that certain tests (e.g. those designed to assess systemic effects) may not be applicable where the presence of leachable materials has been excluded, or where leachables have a known and acceptable toxicity profile;
 - 5) the relationship of device surface area to recipient body size;
 - 6) the existing information based on the literature, experience and non-clinical tests;
 - 7) that the protection of humans is the primary goal of this document, a secondary goal being to ensure animal welfare and to minimize the number and exposure of test animals.
- c) If extracts of the devices are prepared, the solvents and conditions of extraction used shall be appropriate to the nature and use of the final product.
- d) Positive and negative controls shall be used where appropriate.
- e) Test results cannot ensure freedom from potential biological hazard, thus biological investigations shall be followed by careful observations for unexpected adverse reactions or events in humans during clinical use of the device.

A bibliography of International Standards and guidelines on biological-response test methods is given at the end of the text.

5.2 Initial evaluation tests

5.2.1 General

The tests that shall be considered for initial biological response are given in 5.2.2 to 5.2.10.

5.2.2 Cytotoxicity

With the use of cell culture techniques, these tests determine the lysis of cells (cell death), the inhibition of cell growth, and other effects on cells caused by medical devices, materials and/or their extracts. Cytotoxicity tests are described in ISO 10993-5.

5.2.3 Sensitization

These tests estimate, using an appropriate model, the potential of medical devices, materials and/or their extracts for contact sensitization. These tests are appropriate because exposure or contact to even minute amounts of potential leachables can result in allergic or sensitization reactions. Sensitization tests are described in ISO 10993-10.

5.2.4 Irritation

These tests estimate the irritation potential of medical devices, materials and/or their extracts, using appropriate sites for implant tissue such as skin, eye and mucous membrane in a suitable model. The test(s) performed should be appropriate for the route (skin, eye, mucosa) and duration of exposure or contact to

determine irritant effects of devices, materials and potential leachables. Irritation tests are described in ISO 10993-10.

5.2.5 Intracutaneous reactivity

These tests assess the localized reaction of tissue to medical device extracts. These tests are applicable where determination of irritation by dermal or mucosal tests are inappropriate (e.g. medical devices having access to the blood path). These tests may also be useful where extractables are hydrophobic. Intracutaneous reactivity tests are described in ISO 10993-10.

5.2.6 Systemic toxicity (acute toxicity)

These tests estimate the potential harmful effects of either single or multiple exposures, during a period of less than 24 h, to medical devices, materials and/or their extracts in an animal model. These tests are appropriate where contact allows potential absorption of toxic leachables and degradation products.

Pyrogenicity tests are included to detect material-mediated pyrogenic reactions of extracts of medical devices or materials. No single test can differentiate pyrogenic reactions that are material-mediated from those due to endotoxin contamination. Systemic toxicity tests are described in ISO 10993-11.

Immunotoxicity tests should be considered only for devices where data from other sources is suggestive of immunotoxicological effects.

Systemic toxicity tests may be included in subacute and subchronic toxicity test protocols and implantation test protocols.

5.2.7 Subacute and subchronic toxicity

These tests determine the effects of either single or multiple exposures or contact to medical devices, materials and/or their extracts for a period not less than 24 h but not greater than 10 % of the total life-span of the test animal (e.g. up to 90 days in rats). These tests may be waived for materials with chronic toxicity data. The reason for waiving of the tests should be included in the final report. These tests should be appropriate for the route and duration of contact. Subchronic toxicity tests are described in ISO 10993-11.

5.2.8 Genotoxicity

These tests use mammalian or non-mammalian cell culture or other techniques to determine gene mutations, changes in chromosome structure and number, and other DNA or gene toxicities caused by medical devices, materials and/or their extracts. Genotoxicity tests are described in ISO 10993-3.

5.2.9 Implantation

These tests assess the local pathological effects on living tissue, at both the gross level and microscopic level, of a sample of a material or final product that is surgically implanted or placed in an implant site or in a tissue appropriate to the intended application (e.g. special dental usage tests). These tests should be appropriate for the route and duration of contact. For a material, these tests are equivalent to subchronic toxicity tests if systemic effects are also investigated. Implantation tests are described in ISO 10993-6.

Implantation test protocols may be expanded to include systemic toxicity tests, subacute and subchronic toxicity tests, and chronic toxicity tests.

5.2.10 Haemocompatibility

These tests evaluate, using an appropriate model or system, the effects of blood-contacting medical devices or materials on blood or blood components. Specific haemocompatibility tests may also be designed to simulate the geometry, contact conditions and flow dynamics of the device or material during clinical applications.

Haemolysis tests determine the degree of red blood cell lysis and the release of haemoglobin caused by medical devices, materials and/or their extracts *in vitro*. Haemocompatibility tests are described in ISO 10993-4.

5.3 Supplementary evaluation tests

5.3.1 General

The supplementary biological evaluation tests that shall be considered are given in 5.3.2 to 5.3.5.

5.3.2 Chronic toxicity

These tests determine the effects of either single or multiple exposures to medical devices, materials and/or their extracts during at least 10 % of the life-span of the test animal (e.g. more than 90 days in rats). These tests should be appropriate for the route and duration of exposure or contact. Chronic toxicity tests are described in ISO 10993-11.

Chronic toxicity tests may be included in subacute and subchronic toxicity test protocols and implantation test protocols.

5.3.3 Carcinogenicity

These tests determine the tumorigenic potential of medical devices, materials and/or their extracts from either single or multiple exposures or contacts during the major portion of the life-span of the test animal. These tests may be designed in order to examine both chronic toxicity and tumorigenicity in a single experimental study. Carcinogenicity tests should be conducted only if there are suggestive data from other sources. These tests should be appropriate for the route and duration of exposure or contact. Carcinogenicity tests are described in ISO 10993-3.

5.3.4 Reproductive and developmental toxicity

These tests evaluate the potential effects of medical devices, materials and/or their extracts on reproductive function, embryonic development (teratogenicity), and prenatal and early postnatal development. Reproductive/developmental toxicity tests or bioassays should only be conducted when the device has potential impact on the reproductive potential of the subject. The application site of the device should be considered. Reproductive and developmental toxicity tests are described in ISO 10993-3.

5.3.5 Biodegradation

Where the potential for resorption and/or degradation exists, corresponding tests may determine the processes of absorption, distribution, biotransformation and elimination of leachables and degradation products of medical devices, materials and/or their extracts. Biodegradation tests are described in ISO 10993-9.

6 Selection of biological evaluation tests

Evaluation may include both a study of relevant experience and actual testing. Such an evaluation may result in the conclusion that no testing is needed if the material has a demonstrable history of use in a specified role that is equivalent to that of the device under design.

Table 1 identifies the initial evaluation tests that shall be considered for each device and duration category. Table 2 identifies the supplementary evaluation tests that shall be considered for each device and duration category.

Due to the diversity of medical devices, it is recognized that not all tests identified in a category will be necessary or practical for any given device. It is indispensable for testing that each device be considered on its own merits: additional tests not indicated in the table may be necessary.

The tests considered and the rationale for selection and/or waiving of tests shall be recorded.

7 Assurance of test methods

7.1 Test method assurance

The test methods used in the biological evaluation shall be sensitive, precise and accurate. The test results should be reproducible (interlaboratory) as well as repeatable (intralaboratory).

7.2 Continued assurance

The assurance that a material is initially acceptable for its intended use in a medical device, and its continued acceptability in the long term, is an aspect of a quality management system (see A.2, "Subclause 7.2").

NOTE ISO 9001 specifies the requirements for such quality management systems. ISO 9004 provides more detailed guidance for designing and manufacturing products.

Table 1 — Initial evaluation tests for consideration

Medical device categorization by		Biological effect								
Category	Contact	Contact duration (see 4.3) A — Limited (< 24 h) B — prolonged (24 h to 30 days) C — permanent (> 30 days)	Cytotoxicity	Sensitization	Irritation or intracutaneous reactivity	Systemic toxicity (acute)	Subacute and subchronic toxicity	Genotoxicity	Implantation	Haemocompatibility
Surface device	Skin	A	x	x	x					
		B	x	x	x					
		C	x	x	x					
	Mucosal membrane	A	x	x	x					
		B	x	x	x					
		C	x	x	x		x	x		
	Breached or compromised surface	A	x	x	x					
		B	x	x	x					
		C	x	x	x		x	x		
External communicating device	Blood path, indirect	A	x	x	x	x				x
		B	x	x	x	x				x
		C	x	x		x	x	x		x
	Tissue/bone/dentin	A	x	x	x					
		B	x	x	x	x	x	x	x	
		C	x	x	x	x	x	x	x	
	Circulating blood	A	x	x	x	x				x
		B	x	x	x	x	x	x	x	x
		C	x	x	x	x	x	x	x	x
Implant device	Tissue/bone	A	x	x	x					
		B	x	x	x	x	x	x	x	
		C	x	x	x	x	x	x	x	
	Blood	A	x	x	x	x	x		x	x
		B	x	x	x	x	x	x	x	x
		C	x	x	x	x	x	x	x	x
NOTE This table is a framework for the development of an assessment programme and is not a checklist (see Clause 6).										

Table 2 — Supplementary evaluation tests for consideration

Medical device categorization by			Biological effect			
Category	Contact	Contact duration (see 4.3) A — Limited (< 24 h) B — prolonged (24 h to 30 days) C — permanent (> 30 days)	Chronic toxicity	Carcinogenicity	Reproductive/developmental	Biodegradation
Surface device	Skin	A				
		B				
		C				
	Mucosal membrane	A				
		B				
		C				
	Breached or compromised surface	A				
		B				
		C				
External communicating device	Blood path, indirect	A				
		B				
		C	x	x		
	Tissue/bone/dentin	A				
		B				
		C	x	x		
	Circulating blood	A				
		B				
		C	x	x		
Implant device	Tissue/bone	A				
		B				
		C	x	x		
	Blood	A				
		B				
		C	x	x		

NOTE This table is a framework for the development of an assessment programme and is not a checklist (see Clause 6).

Annex A (informative)

Rationale

A.1 General

This part of ISO 10993 is concerned with the safety-in-use of medical devices and materials. It is intended to assess the biological response of devices and materials as part of the overall evaluation and development of devices and materials. Like other current work on this subject, it addresses the determination of the effects of devices and materials on tissue in a general way, rather than for specific individual applications. This part of ISO 10993 thus classifies medical devices into broad categories and indicates, in matrices, the biological tests that are thought to be relevant for consideration for each device category.

Expanded test protocols may combine initial evaluation tests. Examples could be sensitization, irritation and intracutaneous reactivity tests in a single, expanded protocol; or systemic toxicity, subacute and subchronic toxicity, chronic toxicity and implantation tests in a single, expanded protocol.

The range of biological hazards is wide. The tissue interaction of a material cannot be considered in isolation from the overall device design. The best material with respect to tissue interaction may result in a less functional device, tissue interaction being only one characteristic of a material. Where the material is intended to interact with tissue in order for the device to perform its function, evaluation takes on dimensions not generally addressed in standards and guidelines to date.

Biological reactions that are adverse for a material in one application may not be adverse for the use of the material in a different application. Biological testing relies upon animal models and a material cannot, therefore, be conclusively shown to have the same tissue reactions in humans. In addition, differences between humans suggest that some patients may have adverse reactions even to well-established materials.

Currently, biological testing relies on animal models. However, as scientific knowledge advances our understanding of basic mechanisms, preference should be given to *in vitro* models in situations where scientific evidence yields equally relevant information.

For medical devices and materials, the application of a rigid set of test methods and pass/fail criteria might result in either an unnecessary restriction, or a false sense of security, in their use. Where a particular application warrants, experts in the product or application area involved may choose to establish specific tests and criteria specified in a product-specific vertical standard.

This part of ISO 10993 is not, therefore, intended to be a set of definitive statements to be followed by individuals not qualified by training and experience; it should be applied with interpretation and judgement by the appropriate professionals qualified by training and experience, taking into consideration the factors relevant to the device/material, its intended use, and the current knowledge of the device/material provided by scientific literature and previous clinical experience.

A.2 Rationale for specific clauses

The following are rationale for specific clauses in this part of ISO 10993, with clause numbers parallel to those in the body of the document.

Subclause 3.1: It is worth acknowledging that there are important material characteristics other than biological ones that should be taken into account in the overall design of the device.

Subclause 3.6: Evaluation may include both a study of relevant experience and actual testing. Such an evaluation may result in the conclusion that no testing is needed if the material has a demonstrable history of use in a specific role that is the same as that of the device under design.


Evaluation should consider intended and unintended interactions between the material and tissues.

Subclause 3.8: This clause is intended to avoid the need for redundant experiments when information about the material and/or device is available from other sources.

Subclause 7.2: The selection and evaluation of materials that will come into contact with tissue require a structured process so that the materials incorporated in the final design contribute to the overall biological evaluation assurance of the design.

Annex B
(informative)

Flow chart to aid in ensuring a systematic approach to biological evaluation of medical devices



^a Refers to clause/subclause in main text.

Figure B.1

Bibliography

International Organization for Standardization (ISO)

- [1] ISO 7405:1997, *Dentistry — Preclinical evaluation of biocompatibility of medical devices used in dentistry — Test methods for dental materials*
- [2] ISO 9000:2000, *Quality management systems — Fundamentals and vocabulary*
- [3] ISO 9001:2000, *Quality management systems — Requirements*
- [4] ISO 9004:2000, *Quality management systems — Guidelines for performance improvements*
- [5] ISO 10993-3:—¹⁾, *Biological evaluation of medical devices — Part 3: Tests for genotoxicity, carcinogenicity and reproductive toxicity*
- [6] ISO 10993-4:2002, *Biological evaluation of medical devices — Part 4: Selection of tests for interactions with blood*
- [7] ISO 10993-5:1999, *Biological evaluation of medical devices — Part 5: Tests for in vitro cytotoxicity*
- [8] ISO 10993-6:1994, *Biological evaluation of medical devices — Part 6: Tests for local effects after implantation*
- [9] ISO 10993-9:1999, *Biological evaluation of medical devices — Part 9: Framework for identification and quantification of potential degradation products*
- [10] ISO 10993-10:2002, *Biological evaluation of medical devices — Part 10: Tests for irritation and delayed-type hypersensitivity*
- [11] ISO 10993-11:1993, *Biological evaluation of medical devices — Part 11: Tests for systemic toxicity*
- [12] ISO 13485:2003, *Medical devices — Quality systems — System requirements for regulatory purposes*
- [13] ISO 13488:1996, *Quality systems — Medical devices — Particular requirements for the application of ISO 9002*
- [14] ISO/IEC 17025:1999, *General requirements for the competence of testing and calibration laboratories*

Organization for Economic Cooperation and Development (OECD)

- [15] *Guidelines for the testing of chemicals — Section 4: Health effects*

1) To be published. (Revision of ISO 10993-3:1992)

Annex ZA
(informative)

Clauses of this European Standard addressing essential requirements or other provisions of EU Directives

This European Standard has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association and supports essential requirements of EU Directive 93/42/EEC.

WARNING Other requirements and other EU Directives may be applicable to the product(s) falling within the scope of this standard.

Compliance with these clauses of this standard provides one means of conforming with the specific essential requirements of the Directive concerned and associated EFTA regulations.

BSI — British Standards Institution

BSI is the independent national body responsible for preparing British Standards. It presents the UK view on standards in Europe and at the international level. It is incorporated by Royal Charter.

Revisions

British Standards are updated by amendment or revision. Users of British Standards should make sure that they possess the latest amendments or editions.

It is the constant aim of BSI to improve the quality of our products and services. We would be grateful if anyone finding an inaccuracy or ambiguity while using this British Standard would inform the Secretary of the technical committee responsible, the identity of which can be found on the inside front cover. Tel: +44 (0)20 8996 9000. Fax: +44 (0)20 8996 7400.

BSI offers members an individual updating service called PLUS which ensures that subscribers automatically receive the latest editions of standards.

Buying standards

Orders for all BSI, international and foreign standards publications should be addressed to Customer Services. Tel: +44 (0)20 8996 9001. Fax: +44 (0)20 8996 7001. Email: orders@bsi-global.com. Standards are also available from the BSI website at <http://www.bsi-global.com>.

In response to orders for international standards, it is BSI policy to supply the BSI implementation of those that have been published as British Standards, unless otherwise requested.

Information on standards

BSI provides a wide range of information on national, European and international standards through its Library and its Technical Help to Exporters Service. Various BSI electronic information services are also available which give details on all its products and services. Contact the Information Centre. Tel: +44 (0)20 8996 7111. Fax: +44 (0)20 8996 7048. Email: info@bsi-global.com.

Subscribing members of BSI are kept up to date with standards developments and receive substantial discounts on the purchase price of standards. For details of these and other benefits contact Membership Administration. Tel: +44 (0)20 8996 7002. Fax: +44 (0)20 8996 7001. Email: membership@bsi-global.com.

Information regarding online access to British Standards via British Standards Online can be found at <http://www.bsi-global.com/bsonline>.

Further information about BSI is available on the BSI website at <http://www.bsi-global.com>.

Copyright

Copyright subsists in all BSI publications. BSI also holds the copyright, in the UK, of the publications of the international standardization bodies. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI.

This does not preclude the free use, in the course of implementing the standard, of necessary details such as symbols, and size, type or grade designations. If these details are to be used for any other purpose than implementation then the prior written permission of BSI must be obtained.

Details and advice can be obtained from the Copyright & Licensing Manager. Tel: +44 (0)20 8996 7070. Fax: +44 (0)20 8996 7553. Email: copyright@bsi-global.com.

BSI
389 Chiswick High Road
London
W4 4AL